

RA300 Audiogram of OSCAR™ V Electro-Acoustic Ear
(Convenient record of daily biological check)

RA300 AUDIOMETER
TREMETRICS INC.
(C) 1989, 1990,
1995, 1996
DATE 04/17/96
TIME 14 = 29
TEST# 15 - 00014
SUBJECT:

X

SS#/ID#	99999	
CURRENT AUDIOGRAM		
DAILY BIOLOGICAL		
FREQ.	L/DB	R/DB
1KHZ TEST	75	75
500HZ	70	70
1000HZ	75	75
2000HZ	70	70
3000HZ	70	70
4000HZ	70	70
6000HZ	75	80
8000HZ	70	70

MODE PULSED
SERIAL# 0002
VER. 2 REV. A
CAL. ANSI 1989 STD
CAL. DATE 04/96
EXAMINER SS#/ID#
454502101

X

77004-0010
981217

COLL
Health And Safety Inc.
15-145 Traders Blvd. East
Mississauga, ON L4Z 3L3
Ph: (905) 625-3994
Toll Free: (800) 267-9317
Fax: (905) 625-6684
www.collhealth.com