Slides for Aeromedical Model*



Test No. 1, SA-BRL-4
Visual Acuity at Far
With results calibrated at 20/20, 20/30, 20/40, 20/50, 20/70, 20/100 and 20/200. Tests can be administered binocularly or with either eye occluded.



Test No. 2, SSLA-1
Visual Acuity at Far
With results calibrated at 20/15, 20/20, 20/25, 20/30, 20/40, 20/50, 20/70, 20/100 and 20/200. Monocular test only. When test for acuity better than 20/20 is desired.



Test No. 3, SCD-1
Color Perception at Far
Consists of six accurately and authentically reproduced Ishihara Pseudo-Isochromatic plates. Detects color deficiency, but does not classify.



Test No. 4, SLP-1
Lateral Phoria at Far
Measures Esophoria or Exophoria in 1 prism diopter steps.



Test No. 5, SVP-1
Vertical Phoria at Far
Measures Right or Left Hyperphoria in one-half prism diopter steps.



Test No. 6, SA-BRL-2A
Visual Acuity at Near
With results calibrated at 20/20, 20/30, 20/40, 20/50, 20/70 and 20/100. Tests can be administered binocularly or with either eye occluded.

Lenses with instrument:
- Prism lens unit for use with phoria tests to detect deception. Base in prism for SLP-1 test, base up prism for SVP-1 test.
- Intermediate Lens Unit for testing visual acuity at 32 inches.

***The FAA approves Titmus Vision Screeners for testing pilots' vision. The built-in perimeter system is not FAA approved.**

Improvements have been made to test slides to enhance viewing performance. **Patent Pending.**

BACOU USA

TITMUS®